

1 - ブロモブタンのマウスを用いた
吸入によるがん原性試験報告書

試験番号 : 0561

APPENDICES

APPENDICES

- APPENDIX 1-1 IDENTITY OF 1 - BROMOBUTANE IN THE 2-YEAR INHALATION STUDY
- APPENDIX 1-2 STABILITY OF 1 - BROMOBUTANE IN THE 2-YEAR INHALATION STUDY
- APPENDIX 2 ENVIRONMENTAL CONDITIONS OF INHALATION CHAMBER IN THE 2-YEAR INHALATION STUDY OF 1 - BROMOBUTANE
- APPENDIX 3 METHODS, UNITS AND DECIMAL PLACE FOR HEMATOLOGY AND BIOCHEMISTRY IN THE 2-YEAR INHALATION STUDY OF 1 - BROMOBUTANE

APPENDIX 1-1

IDENTITY OF 1 - BROMOBUTANE
IN THE 2-YEAR INHALATION STUDY

IDENTITY OF 1-BROMOBUTANE IN THE 2-YEAR INHALATION STUDY

Test Substance : 1-Bromobutane (Wako Pure Chemical Industries, Ltd.)

A. Lot No. : KLG0007



1. Spectral Data

Mass Spectrometry

Instrument : Hitachi M-80B Mass Spectrometer

Ionization : EI (Electron Ionization)

Ionization Voltage : 70eV




Result: The mass spectrum was consistent with literature spectrum.

(*McLafferty FW, ed. 1994. Wiley Registry of Mass Spectral Data. 6th ed. New York, NY:John Wiley and Sons.)


Infrared Spectrometry

Instrument : Shimadzu FTIR-8200PC Infrared Spectrometer

Cell : KBr Liquid Cell

Resolution : 4 cm^{-1} 

Infrared Spectrum of Test Substance



Infrared Spectrum of Literature Data*

Result: The infrared spectrum was consistent with literature spectrum.
(*Performed by Wako Pure Chemical Industries, Ltd.)

2. Conclusion: The test substance was identified as 1-bromobutane by mass spectrum and infrared spectrum.

B. Lot No. : EWL0012


1. Spectral Data

Mass Spectrometry


Instrument : Hitachi M-80B Mass Spectrometer

Ionization : EI (Electron Ionization)

Ionization Voltage : 70eV



Mass Spectrum of Test Substance



Mass Spectrum of Literature Data*


Result: The mass spectrum was consistent with literature spectrum.

(*McLafferty FW, ed. 1994. Wiley Registry of Mass Spectral Data. 6th ed. New York, NY:John Wiley and Sons.)


Infrared Spectrometry

Instrument : Shimadzu FTIR-8200PC Infrared Spectrometer

Cell : KBr Liquid Cell

Resolution : 4 cm^{-1} 

Infrared Spectrum of Test Substance



Infrared Spectrum of Literature Data*

Result: The infrared spectrum was consistent with literature spectrum.

(*Performed by Wako Pure Chemical Industries, Ltd.)

2. Conclusion: The test substance was identified as 1-bromobutane by mass spectrum and infrared spectrum.

C. Lot No. : DPN0021


1. Spectral Data

Mass Spectrometry


Instrument : Hitachi M-80B Mass Spectrometer

Ionization : EI (Electron Ionization)

Ionization Voltage : 70eV



Mass Spectrum of Test Substance



Mass Spectrum of Literature Data*


Result: The mass spectrum was consistent with literature spectrum.

(*McLafferty FW, ed. 1994. Wiley Registry of Mass Spectral Data. 6th ed. New York, NY:John Wiley and Sons.)


Infrared Spectrometry

Instrument : Shimadzu FTIR-8200PC Infrared Spectrometer

Cell : KBr Liquid Cell

Resolution : 4 cm^{-1} 

Infrared Spectrum of Test Substance



Infrared Spectrum of Literature Data*

Result: The infrared spectrum was consistent with literature spectrum.

(*Performed by Wako Pure Chemical Industries, Ltd.)

2. Conclusion: The test substance was identified as 1-bromobutane by mass spectrum and infrared spectrum.

APPENDIX 1-2

STABILITY OF 1 - BROMOBUTANE
IN THE 2-YEAR INHALATION STUDY

STABILITY OF 1-BROMOBUTANE IN THE 2-YEAR INHALATION STUDY

Test Substance : 1-Bromobutane (Wako Pure Chemical Industries, Ltd.)

A. Lot No. : KLG0007

1. Gas Chromatography

Instrument : Agilent Technologies 5890A Gas Chromatograph

Column : Methyl Silicone (0.53 mm ϕ \times 60 m)

Column Temperature : 100° C

Flow Rate : 15 mL/min

Detector : FID (Flame Ionization Detector)

Injection Volume : 1 μ L

| Date (date analyzed) | Peak No. | Retention Time (min) | Area (%) |
|-------------------------|----------|-------------------------|-------------|
| 2004.10.15 | 1 | 4.128 | 100 |
| 2005.09.06 | 1 | 4.128 | 100 |

Result: Gas chromatography indicated one major peak (peak No.1) analyzed on 2004.10.15 and one major peak (peak No.1) analyzed on 2005.9.6. No new trace impurity peak in the test substance analyzed on 2005.9.6 was detected.

2. Conclusion: The test substance was stable for the period that the test substance had been used for the study.

B. Lot No. : EWL0012

1. Gas Chromatography

Instrument : Agilent Technologies 5890A Gas Chromatograph

Column : Methyl Silicone (0.53 mm ϕ \times 60 m)

Column Temperature : 100° C

Flow Rate : 15 mL/min

Detector : FID (Flame Ionization Detector)

Injection Volume : 1 μ L

| Date (date analyzed) | Peak No. | Retention Time (min) | Area (%) |
|-------------------------|----------|-------------------------|-------------|
| 2005.09.01 | 1 | 4.130 | 100 |
| 2006.05.24 | 1 | 4.126 | 100 |

Result: Gas chromatography indicated one major peak (peak No.1) analyzed on 2005.9.1 and one major peak (peak No.1) analyzed on 2006.5.24. No new trace impurity peak in the test substance analyzed on 2006.5.24 was detected.

2. Conclusion: The test substance was stable for the period that the test substance had been used for the study.

C. Lot No. : DPN0021

1. Gas Chromatography

Instrument : Agilent Technologies 5890A Gas Chromatograph

Column : Methyl Silicone (0.53 mm ϕ \times 60 m)

Column Temperature : 100° C

Flow Rate : 15 mL/min

Detector : FID (Flame Ionization Detector)

Injection Volume : 1 μ L

| Date (date analyzed) | Peak No. | Retention Time (min) | Area (%) |
|-------------------------|----------|-------------------------|-------------|
| 2006.05.22 | 1 | 4.125 | 100 |
| 2006.11.22 | 1 | 4.126 | 100 |

Result: Gas chromatography indicated one major peak (peak No.1) analyzed on 2006.5.22 and one major peak (peak No.1) analyzed on 2006.11.22. No new trace impurity peak in the test substance analyzed on 2006.11.22 was detected.

2. Conclusion: The test substance was stable for the period that the test substance had been used for the study.

APPENDIX 2

ENVIRONMENTAL CONDITIONS OF INHALATION CHAMBER IN THE 2-YEAR INHALATION STUDY OF 1 - BROMOBUTANE

ENVIRONMENTAL CONDITIONS OF INHALATION CHAMBER IN THE 2-YEAR
INHALATION STUDY OF 1-BROMOBUTANE

| Group Name | Temperature (°C) Mean ± S.D. | Humidity (%) Mean ± S.D. | Ventilation Rate (L/min) Mean ± S.D. | Air Change (time/h) Mean |
|------------|------------------------------------|--------------------------------|--------------------------------------------|--------------------------------|
| Control | 23.1 ± 0.1 | 54.6 ± 1.1 | 782.1 ± 3.9 | 12.0 |
| 20 ppm | 23.1 ± 0.0 | 54.2 ± 1.2 | 782.3 ± 3.3 | 12.0 |
| 50 ppm | 23.0 ± 0.0 | 54.4 ± 1.1 | 782.7 ± 3.5 | 12.0 |
| 125 ppm | 23.0 ± 0.1 | 53.6 ± 1.0 | 783.0 ± 3.7 | 12.0 |

APPENDIX 3

METHODS, UNITS AND DECIMAL PLACE FOR
HEMATOLOGY AND BIOCHEMISTRY IN THE 2-YEAR
INHALATION STUDY OF 1 - BROMOBUTANE

METHODS, UNITS AND DECIMAL PLACE FOR HEMATOLOGY AND BIOCHEMISTRY
IN THE 2-YEAR INHALATION STUDY OF 1 - BROMOBUTANE

| Item | Method | Unit | Decimal place |
|----------------------------------------------------|-------------------------------------------------------------------|---------------------------|---------------|
| Hematology | | | |
| Red blood cell (RBC) | Light scattering method ¹⁾ | $\times 10^6/\mu\text{L}$ | 2 |
| Hemoglobin(Hgb) | Cyanmethemoglobin method ¹⁾ | g/dL | 1 |
| Hematocrit(Hct) | Calculated as $\text{RBC} \times \text{MCV}/10$ ¹⁾ | % | 1 |
| Mean corpuscular volume(MCV) | Light scattering method ¹⁾ | fL | 1 |
| Mean corpuscular hemoglobin(MCH) | Calculated as $\text{Hgb}/\text{RBC} \times 10$ ¹⁾ | pg | 1 |
| Mean corpuscular hemoglobin concentration (MCHC) | Calculated as $\text{Hgb}/\text{Hct} \times 100$ ¹⁾ | g/dL | 1 |
| Platelet | Light scattering method ¹⁾ | $\times 10^3/\mu\text{L}$ | 0 |
| Reticulocyte | Light scattering method ¹⁾ | % | 1 |
| White blood cell(WBC) | Light scattering method ¹⁾ | $\times 10^3/\mu\text{L}$ | 2 |
| Differential WBC | Pattern recognition method ²⁾ (Wright staining) | % | 0 |
| Biochemistry | | | |
| Total protein(TP) | Biuret method ³⁾ | g/dL | 1 |
| Albumin (Alb) | BCG method ³⁾ | g/dL | 1 |
| A/G ratio | Calculated as $\text{Alb}/(\text{TP} - \text{Alb})$ ³⁾ | - | 1 |
| T-bilirubin | Azobilirubin method ³⁾ | mg/dL | 2 |
| Glucose | GlcK·G-6-PDH method ³⁾ | mg/dL | 0 |
| T-cholesterol | CE·COD·POD method ³⁾ | mg/dL | 0 |
| Triglyceride | MGLP·GK·GPO·POD method ³⁾ | mg/dL | 0 |
| Phospholipid | PLD·ChOD·POD method ³⁾ | mg/dL | 0 |
| Aspartate aminotransferase (AST) | JSCC method ³⁾ | IU/L | 0 |
| Alanine aminotransferase (ALT) | JSCC method ³⁾ | IU/L | 0 |
| Lactate dehydrogenase (LDH) | SFBC method ³⁾ | IU/L | 0 |
| Alkaline phosphatase (ALP) | GSCC method ³⁾ | IU/L | 0 |
| γ -Glutamyl transpeptidase (γ -GTP) | JSCC method ³⁾ | IU/L | 0 |
| Creatine kinase (CK) | JSCC method ³⁾ | IU/L | 0 |
| Urea nitrogen | Urease·GLDH method ³⁾ | mg/dL | 1 |
| Sodium | Ion selective electrode method ³⁾ | mEq/L | 0 |
| Potassium | Ion selective electrode method ³⁾ | mEq/L | 1 |
| Chloride | Ion selective electrode method ³⁾ | mEq/L | 0 |
| Calcium | OCPC method ³⁾ | mg/dL | 1 |
| Inorganic phosphorus | PNP·XOD·POD method ³⁾ | mg/dL | 1 |

1) Automatic blood cell analyzer (ADVIA120 : Siemens Medical Solutions Diagnostics)

2) Automatic blood cell differential analyzer (MICROX HEG-120NA : OMRON Corporation)

3) Automatic analyzer (Hitachi 7080 : Hitachi,Ltd.)