アリルアルコールのラットを用いた吸入による2週間毒性試験報告書

試験番号:0885

APPENDICES

APPENDICES

| APPENDIX 1-1 | IDENTITY OF ALLYL ALCOHOL IN THE 2-WEEK INHALATION STUDY |
|--------------|--|
| APPENDIX 1-2 | STABILITY OF ALLYL ALCOHOL IN THE 2-WEEK INHALATION STUDY |
| APPENDIX 2 | ENVIRONMENTAL CONDITIONS OF INHALATION CHAMBER IN THE 2-WEEK INHALATION STUDY OF ALLYL ALCOHOL |
| APPENDIX 3 | METHODS, UNITS AND DECIMAL PLACE FOR HEMATOLOGY AND BIOCHEMISTRY IN THE 2-WEEK INHALATION STUDY OF ALLYL ALCOHOL |

APPENDIX 1 - 1

IDENTITY OF ALLYL ALCOHOL IN THE 2-WEEK INHALATION STUDY

IDENTITY OF ALLYL ALCOHOL IN THE 2-WEEK INHALATION STUDY

Test Substance

: Allyl alcohol (Wako Pure Chemical Industries, Ltd.)

Lot No.

: TWR6089

1. Mass Spectrometry

Instrument


: Hitachi M-80B Mass Spectrometer


Ionization

: EI (Electron Ionization)

Ionization Voltage

: 70eV





Mass Spectrum of Literature Data*

Result: The mass spectrum was consistent with literature spectrum.

(*McLafferty FW, ed. 1994. Wiley Registry of Mass Spectral Data. 6th ed. New York, NY:John Wiley and Sons.)

2. Conclusion: The test substance was identified as allyl alcohol by mass spectrum.

APPENDIX 1 - 2

STABILITY OF ALLYL ALCOHOL IN THE 2-WEEK INHALATION STUDY

STABILITY OF ALLYL ALCOHOL IN THE 2-WEEK INHALATION STUDY

Test Substance

: Allyl alcohol (Wako Pure Chemical Industries, Ltd.)

Lot No.

: TWR6089

1. Gas Chromatography

Instrument

: Agilent Technologies 5890A Gas Chromatograph

Column

: INNOWAX (0.53 mm ϕ × 60 m)

Column Temperature: 100 °C

Flow Rate

: 5 mL/min

Detector

: FID (Flame Ionization Detector)

Injection Volume

 $: 1 \mu L$

| Date Analyzed | Peak No. | Retention Time (min) | Area (%) |
|---------------|----------|-------------------------|-------------|
| 2017.03.06 | 1 | 4.839 | 100 |
| 2017.03.23 | 1 | 4.841 | 100 |

Result: Gas chromatography indicated one major peak (peak No.1) analyzed on 2017.3.6 and one major peak (peak No.1) analyzed on 2017.3.23. No new trace impurity peak in the test substance analyzed on 2017.3.23 was detected.

2. Conclusion: The test substance was stable for the period that the test substance had been used for the study.

APPENDIX 2

ENVIRONMENTAL CONDITIONS OF INHALATION CHAMBER IN THE 2-WEEK INHALATION STUDY OF ALLYL ALCOHOL

ENVIRONMENTAL CONDITIONS OF INHALATION CHAMBER IN THE 2-WEEK INHALATION STUDY OF ALLYL ALCOHOL

| Group Name | Temperature ($^{\circ}$ C) Mean \pm S.D. | Humidity $(\%)$ Mean \pm S.D. | Ventilation Rate (L/min) Mean \pm S.D. | Air Change (time/h) Mean |
|---------------------|---|---------------------------------|--|--------------------------------|
| Control | 22.9 ± 0.1 | 57.2 ± 0.3 | 214.0 ± 0.5 | 12.1 |
| 6.3 ppm | 22.9 ± 0.1 | 57.0 ± 0.4 | 213.4 ± 0.8 | 12.1 |
| $12.5~\mathrm{ppm}$ | 23.0 ± 0.1 | 56.5 ± 0.4 | 214.0 ± 0.5 | 12.1 |
| 25~ m ppm | 22.9 ± 0.2 | 56.9 ± 0.4 | 214.1 ± 0.5 | 12.1 |
| 50~ m ppm | 23.1 ± 0.1 | 54.5 ± 0.5 | 213.7 ± 0.6 | 12.1 |
| 100 ppm | 22.9 ± 0.1 | 54.4 ± 0.7 | 213.8 ± 0.6 | 12.1 |

APPENDIX 3

METHODS, UNITS AND DECIMAL PLACE FOR
HEMATOLOGY AND BIOCHEMISTRY IN THE 2-WEEK
INHALATION STUDY OF ALLYL ALCOHOL

METHODS, UNITS AND DECIMAL PLACE FOR HEMATOLOGY AND BIOCHEMISTRY IN THE 2-WEEK INHALATION STUDY OF ALLYL ALCOHOL

| Item | Method | Unit | Decimal |
|---|--|----------------------------|---------|
| | | | place |
| Hematology | | | |
| Red blood cell (RBC) | Light scattering method 1) | $	imes 10^6/\mu\mathrm{L}$ | 2 |
| Hemoglobin(Hgb) | Cyanmethemoglobin method 1) | g/dL | 1 |
| Hematocrit(Hct) | Calculated as RBC×MCV/10 1) | % | 1 |
| Mean corpuscular volume(MCV) | Light scattering method 1) | fL | 1 |
| Mean corpuscular hemoglobin(MCH) | Calculated as Hgb/RBC×10 10 | pg | 1 |
| Mean corpuscular hemoglobin concentration | Calculated as Hgb/Hct×100 10 | g/dL | 1 |
| (MCHC) | - | | |
| Platelet | Light scattering method 1) | $	imes 10^3 / \mu$ L | 0 |
| Reticulocyte | Light scattering method 1) | % | 1 |
| White blood cell(WBC) | Light scattering method 1) | $	imes 10^3/\mu~ m L$ | 2 |
| Differential WBC | Light scattering method 1) | % | 0 |
| Biochemistry | | | |
| Total protein(TP) | Biuret method 2) | g/dL | 1 |
| Albumin (Alb) | BCG method 2) | g/dL | 1 |
| A/G ratio | Calculated as Alb/(TP-Alb) 2) | _ | 1 |
| T-bilirubin | BOD method 2) | mg/dL | 2 |
| Glucose | GlcK·G-6-PDH method 2) | mg/dL | 0 |
| T-cholesterol | CE·COD·POD method 2) | mg/dL | 0 |
| Triglyceride | MGLP·GK·GPO·POD method 2) | mg/dL | 0 |
| Phospholipid | PLD·ChOD·POD method 2) | mg/dL | 0 |
| Aspartate aminotransferase (AST) | JSCC method ²⁾ | U/L | 0 |
| Alanine aminotransferase (ALT) | JSCC method 2) | U/L | 0 |
| Lactate dehydrogenase (LDH) | JSCC method ²⁾ | U/L | 0 |
| Alkaline phosphatase (ALP) | JSCC method ²⁾ | U/L | 0 |
| γ-Glutamyl transpeptidase (γ-GTP) | JSCC method 2) | U/L | 1 |
| Creatine kinase (CK) | JSCC method ²⁾ | U/L | 0 |
| Urea nitrogen | Urease · GLDH method 2) | mg/dL | 1 |
| Creatinine | Creatinase·SOD·POD method 2) | mg/dL | 2 |
| Sodium | Ion selective electrode method ²⁾ | mEq/L | 0 |
| Potassium | Ion selective electrode method 2) | mEq/L | 1 |
| Chloride | Ion selective electrode method ²⁾ | mEq/L | 0 |
| Calcium | OCPC method 2) | mg/dL | 1 |
| Inorganic phosphorus | PNP·XOD·POD method 2) | mg/dL | 1 |

¹⁾ Automatic blood cell analyzer (ADVIA120 : Siemens Healthcare Diagnostics Inc.)

²⁾ Automatic analyzer (Hitachi 7080 : Hitachi, Ltd.)